

Customer Site
 Datashredders Limited
 Unit 2, Eastwood Industrial Estate,
 Eastwood End,
 Wimblington,
 Cambridgeshire, PE15 0QH.

Certificate Number: 4761
 Date: 17th March 2026

Undertaken by:
 Julian Braddock

Platform Type:
 16m 6400s

Number of Load Points:
 6

Indicator Type:
 Gi650

Serial No.:
 002038

Maximum Capacity:
 50,000kg

Minimum Capacity (20xe):
 400kg

Increment (e):
 20kg

Class:
 III

ECCENTRIC LOADING

Accuracy at load points to be tested:

$$\frac{\text{Max. Capacity}}{\text{No./load points} - 1} = \text{Nearest practical weight}$$

$$\frac{50,000\text{kg}}{(5)} = 10,000 \text{ KG}$$

PASS FAIL

ACCURACY ACROSS LOAD POINTS

(Expanded Load Points) (expanded display)

0e	0e	0e
0e	0e	0e

WEIGH OFFICE

ACCURACY OF INSTRUMENT - Linearity & Hysteresis (expanded display)

Known wt.	Increasing Reading	Decreasing Reading	Error	
400kg	400kg	400kg	0e	
10,000kg	10,000kg	10,000kg	0e	
21,400kg	21,400kg	21,400kg	0e	<input checked="" type="checkbox"/> PASS
43,922kg	43,918kg	43,918kg	-.2e	<input type="checkbox"/> FAIL
48,624kg	48,620kg		-.2e	

ROLLING TESTS Load 43,918 Kg (expanded display)

END	MIDDLE	END	<input checked="" type="checkbox"/> PASS	END	MIDDLE	END
43,918kg	43,918kg	43,918kg	<input type="checkbox"/> FAIL	43,918kg	43,918kg	43,918kg

Deviations recorded as fractions of (e) unless expanded display used and recorded in KG

	TEST	Verification
Permissible error:	0 ≤ m ≤ 500e	+/- 1e
	500 ≤ m ≤ 2000e	+/- 2e
	2000 ≤ m ≤ 10000e	+/- 3e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

No. TR13273
 Dated 13th March 2026

**DECLARACIÓN UE DE CONFORMIDAD**
DECLARATION UE OF CONFORMITY
DÉCLARATION UE DE CONFORMITÉ
Nº 26-1810

Nosotros/We/Nous:

GIROPES, S.L.
Pol. Ind. Empordà Internacional, c/ Molló 3.
E-17469 VILAMALLA (Girona)
-SPAIN-Declaramos bajo nuestra responsabilidad que el producto:
Declare under our responsibility that the product:
Nous déclarons sous notre responsabilité que le produit:Instrumento de pesaje de funcionamiento no automático (IPFNA)
The non-automatic weighing instrument (NAWI)
Instrument de pesage à fonctionnement non automatique

Model / Model / Modèle:

Gi650

Nº de serie / Serial number / N° Série

26030364Al cual se refiere la presente declaración, es conforme a:
To which declaration refers to, conform with:
A la quelle se refaire la présente déclaration, et conforme aux:

Marcado CE Mark applied CE marquage	Directiva EU EU Directive EU Directive	Normas Standards Normes	Homologaciones Type approvals Approbation du modèle
0122 (1)	2014/31/EU	EN 45501	0200-NAWI-07060 FORCE (0200)
	2014/30/EU	EN 55022:2010+AC:2011	
	2014/35/EU	EN 60950-1/A12:2011	

(1) El organismo notificado NMI Certin BV (0122) llevó a cabo una auditoría para el módulo D de la directiva NAWI y emitió el certificado CE-258 para GIROPES. S.L.

(1) The notified body NMI Certin BV (0122) performed Module D audit for NAWI directive and issued the certificate CE-258 for GIROPES S.L.

(1) L'organisme notifié NMI Certin BV (0122) a examiné le respect de la Directive NAWI par le Module D et a émis à GIROPES S.L.. le certificat CE-258.

La verificación CE es válida para el siguiente lugar de instalación/localización/área de uso:

The EC-verification is valid for the following place of instalation/location/area of use:

La vérification CE est valide pour l'emplacement/l'endroit d'utilisation/sphère d'utilisation suivant:

Vilamalla, 06/03/2026

Datashredders Limited
PE15 0QH - Wimblington - Cambridgeshire
UNITED KINGDOM (g=9,8129 m/s²)
GIROPES, S.L.Jordi Ribalta.
Technical Manager