



Data Shredders Ltd  
44 Hostmoor Avenue  
March  
Cambridgeshire  
PE15 OAX

**Certificate of Accuracy** 4491

Date: 3-12-24

Platform Type :	64005	No. of load points :	6
Indicator Type :	3590		
Serial No. :	04515020		
Maximum Cap. :	50000kg	Minimum Capacity (20 x e) :	400kg
Increment (e) :	20kg	Class :	III

**ECCENTRIC LOADING** Accuracy across load points (Expanded Display )

Accuracy at load points to be tested:  
Max. Capacity = Nearest  
No./load points - 1 practical weight

50000 = 10000 KG  
(6-1) Pass  Fail

✓	✓	✓
✓	✓	✓

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display )

Known wt.	Increasing reading	Decreasing reading	Error
400	400	400	0
11400	11400	11400	0
22400	22400	22400	-20
33860	33820	33860	-40
43860	43820		-40

Pass

Fail

**ROLLING TESTS** LOAD: 43820 kg (Expanded display )

END	MIDDLE	END	Pass <input checked="" type="checkbox"/>	END	MIDDLE	END
43820	43820	43820	Fail <input type="checkbox"/>	43820	43820	43820

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	Verification
Permissible errors:	0 ≤ m ≤ 500e	+/-1e
	500 ≤ m ≤ 2000e	+/-0.5e
	2000 ≤ m ≤ 10000e	+/-1e
		+/-1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by M Bernisford

No. TR12336 dated 14-3-24